

Certificate of Analysis

Product: Water Purified
Item Number: WAU
Grade: USP
Lot Number: 5001191/1.3
Manufacture Date: 03/10/2025
Expiration Date: 09/09/2026
Country of Origin: United States

Tested Property	Specification	Analysis
Appearance	Clear, colorless liquid	Passes
Water Conductivity (Point of fill)	$\leq 1.3 \mu\text{S/cm}$ at 25°C	0.6
Microbial Limits	$\leq 100 \text{ CFU/mL}$	<1
pH	5.0 - 7.0	6.3
Substances Reducing Permanganate	The red color does not disappear	Passes
Total Organic Carbon	$< 0.50 \text{ mg C/L}$	0.01

"Point of Fill" refers to the testing of conductivity in the production loop before the water is filled into final packaging.

[Water Purified](#)

Note: The information and recommendations of Lab Alley concerning this product are based upon laboratory tests and experience. To the best of our knowledge and belief these are true and accurate, however Lab Alley assumes no obligation or liability for the information in this document. Since conditions of actual use are beyond our control, any recommendations or suggestions regarding merchantability and fitness for particular purposes are made without warranty, expressed or implied.

Phone: 512-668-9918, Fax: 512-886-4008,
E-mail: customercare@laballey.com,
www.laballey.com.

12501 Pauls Valley Road, Suite A, Austin, Texas 78737.

© copyright: 2025 Lab Alley